



DIAGNOSTIC AUTOMATION, INC.

21250 Califa St. Suite 102 and 116, Woodland Hills, CA 91367 USA

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com

Material Safety Data Sheet

ASO

Cat # 330-50

(Serology Test Kit)

IDENTIFICATION OF THE PREPARATION AND COMPANY NAME.

Supplier: Diagnostic Automation Inc.
21250 Califa Street, Suite 102 and 116,
Woodland Hills, CA 91367 USA

Emergency Telephone number: Please contact the local hospitals.

Product Name: ASO TEST KIT
Catalog No: 330-50

Composition

CHEMICAL NAME	%	CAS#
Buffered Salts, Antibodies	Mixture	
Other components either non-hazardous or at concentrations below that requiring hazardous listing.		

Physical Data

Boiling Point	Melting Point	Specific Gravity
100°C	N/A	1 (Water = 1)
Vapor Pressure	Percent Volatile by Vol.	Vapor Density
N/A	> 90	0.6 (AIR = 1)
Evaporation Rate	Solubility in Water	Appearance/Odor
1 (WATER = 1)	Soluble	RPR antigen – gray color Positive & Negative Control - clear

Fire and Explosive Hazard Data

Flash Point (Method Used)	Flammable Limits LEL UEL
Extinguishing Media	
Use media appropriate for surrounding materials.	
Special Fire Fighting Procedures	
Explosion Potential	
Reagents containing Sodium Azide may combine with copper and lead plumbing to form highly explosive metal azide.	

Health and Hazard Data

Threshold Limit Value (TLV)
Effects of Overexposure
Positive and Negative controls predated using human sera found negative for hepatitis B surface antigen (HBsAg) and HIV-III by FDA required test; however, handle as if potentially infectious.
Emergency and First Aid Procedures

Reactivity Data

Stability: Unstable <input type="checkbox"/> Stable <input checked="" type="checkbox"/>	Polymerization: May Occur <input type="checkbox"/> Will Not Occur <input checked="" type="checkbox"/>
Incompatibility (Materials to avoid)	
N/A	
Decomposition Products	
N/A	
Conditions to be avoided	
N/A	

Spill and Disposal Procedures

Steps to be taken in case of Leaks or Spills
Flush area with large amounts of water, clean with detergent until all reminiscence are gone.
Waste Disposal Method
Follow Federal, State and local regulations.

Personal Protection Information

Ventilation	Mechanical
Good General Ventilation	N/A
	Local Exhaust
	N/A
Eye Protection	Protective Gloves
Goggles with side shields recommended.	Gloves Required
Other Protective Equipment	
Laboratory Cost: Eyewash: Emergency Shower.	
Measures to be taken during repair and maintenance equipment that has been in contact with the material	
Wash equipment with water. Dispose as in VII above	

Special Precautions

Precautions to be taken in Handling and Storage

Store at temperature listed on container.

Other Precautions

Do not pipette by mouth. Avoid contact or ingestion.

Other Information

This Product is labeled in accordance with CFR21 (Food and Drugs), Section 809.10.

The information contained herein has been compiled from data presented in various technical sources believed to be accurate. Diagnostic Automation, Inc. makes no warranties and assumes no liability in connection with the use of this information. It is the user's responsibility to determine the suitability of this information and to assure the adoption of necessary safety precautions.

Date Adopted	2016-06-10
REF 330-50	DA-MONO



DIAGNOSTIC AUTOMATION, INC.

21250 Califa St. Suite 102 and 116, Woodland Hills, CA 91367 USA

Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision Date: 2003-02-12