



CORTEZ DIAGNOSTICS, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, CA 91367

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com

MATERIAL SAFETY DATA SHEET

OneStep

FSH Urine

RapiCard™ InstaTest

Manufacturer: Diagnostic Automation Inc.
21250 Califa Street ,Suite 102 and 116
Woodland Hills, CA 91367 USA

Product Name: FSH Urine RapiCard
Catalog No. 198220-1

Composition: Test Device:
Quantity:
Regulatory Information:

Sodium Azide
>0.0001%
OMSHA PEL 0.3 mg/cubic meter CL
ACGIH,TLV: TWA Ceiling Limit 0.1 ppm

Physical Data

Boiling Point:	Not Applicable
Melting Point:	Decomposition at 300 °C
Vapor Pressure (mm Hg):	Not Applicable
Maximum Vapor Conc. (g/m ³):	Not Applicable
Density (D204):	Not Applicable
Appearance:	Off-white test device
Specific Gravity (Water at 1.0)	1.850 (Sodium Azide)
Evaporation Rate:	Not Applicable
Odor:	None Known

Hazards Identification

Medical Condition Caused By Exposure:

Primary Route(s) of Entry:

First Aid:

Eyes/Skin Hypersensitivity

Eyes and Skin Contact, Ingestion, Inhalation.

In all cases call physician or poison control center.

Sodium Azide: Immediately flush eyes or skin with copious amounts of water for at least 15 minutes while removing contaminated clothing and shoes

Assure adequate flushing of the eyes by separating the eyelids with fingers.

If inhaled: removed to fresh air. If not breathing give artificial respiration. If breathing is difficult, give oxygen.

If swallowed wash out mouth with water

provided person is conscious. Washing contaminated clothing before re-use. Discard contaminated shoes.

Reactivity Data

Stability:

Stable

Reactivity:

Non-reactive within this kit

Conditions to Avoid:

None Known

Incompatibility:

None Known

Materials to Avoid:

Aromatic Hydrocarbons

Hazardous Polymerization:

Will not occur

Hazardous Decomposition Products: None Known

Environmental Protection Procedures

Spill Response Procedures:

Wear appropriate personal protective equipment, including but not limited to protective eyewear, clothing and gloves. Clean up with soap and water.

Waste Disposal Method:

Normal disposal container.

Liquids Disposal:

Wash down drain and flush with excess water.

First Aid Measures

Eye Protection:

Wear safety glasses or full face shield

Skin Protection:

Wear protective clothing and gloves, remove all clothing if contaminated

Respiratory:

None required

Ventilation:

Normal

Other Precautions:

Not needed

Fire Fighting Measures

Flash Point:

Not Applicable

Auto Ignition Temperature:

Not Applicable

LEL:

Not Applicable

UEL:

Not Applicable

Extinguishing Media:

Sodium Azide: Dry Chemical Powder

Unusual Fire & Explosive Hazards:

None Known

Special Fire-Fighting Procedures:

Non Flammable

Special Precautions

Hygienic Practices in Handling
And Storage:

Storage conditions per package insert instructions.

Precautions for Repair & Maintenance
Of Contaminated Material:

Not Applicable

Other Precautions:

For In Vitro Diagnostic Use Only. Do not use internally. Do not apply to eyes or skin, ingest or inhale. Do not mix reagents from different lots.

V: R- & S-Phrases

R28: Very toxic if swallowed

R32: Contact with acid liberates very toxic gases

S28: After contact with skin, wash immediately with plenty of water

Other Information

For further information, please contact Cortez Diagnostics Inc.

Date Adopted	2015-10-01
REF 198220-1	DA-LH 2015



CORTEZ DIAGNOSTICS, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, CA 91367

Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision B Date: 2015-10-01