



DIAGNOSTIC AUTOMATION, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, CA 91367

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com

IVD



See external label



96 tests

REF

1403-11

Herpes Simplex Virus 2 IgG

HSV 2 IgG

REF 1403-11

Test	HSV 2 IgG ELISA
Method	ELISA: Enzyme Linked Immunosorbent Assay
Principle	Indirect: Antigen Coated Plate
Detection Range	Qualitative Positive; Negative control & Cut off
Sample	5ul Serum
Total Time	~75 min
Shelf Life	12 Months from the manufacturing date

** Laboratory results can never be the only base of a medical report. The patient history and further tests have to be taken into account*

NAME AND INTENDED USE

Diagnostic Automation/ Cortez Diagnostics, Inc.
21250 Califa St, Suite 102 and 116, Woodland Hills, CA 91367 USA Phone: 818-591-3030, Fax : 818-591-8383
Email: onestep@rapidtest.com Website: www.rapidtest.com

The DIAGNOSTIC AUTOMATION ELISA, HSV 2 IgG is intended for use in evaluating a patient's serologic status to herpes simplex virus (HSV) infection, or for evaluating paired sera for the presence of a significant increase in herpes specific IgG.

SUMMARY AND EXPLANATION OF THE TEST

Herpes Simplex Virus is a common pathogen and its primary infection is usually asymptomatic. There are two immunologically distinct types of HSV: Type 1 and Type 2. HSV 1 is generally associated with oral infection and lesions above the waist, and HSV 2 is associated with genital infections and lesions below the waist. Clinical cases primarily are 1) eczema herpeticum with eczematous skin changes with numerous lesions, 2) Gingivo-stomatitis and 3) Herpes sepsis, almost only found in newly born or premature infants. MICROWELL ELISA HSV IgG is an accurate serologic method to detect HSV specific antibody in serum sample.

PRINCIPLE OF THE TEST

Purified HSV antigen is coated on the surface of microwells. Diluted patient serum is added to wells, and the HSV 2 IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away. After adding enzyme conjugate, it binds to the antibody-antigen complex. Excess enzyme conjugate is washed off and TMB Chromogenic substrate is added. The enzyme conjugate catalytic reaction is stopped at a specific time. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample. The results are read by a microwell reader compared in a parallel manner with calibrator and controls.

MATERIALS PROVIDED

1. Microwell Strips: Purified HSV 2 antigen coated virus	12 X 8 wells
2. Sample Diluent: Blue Color Solution	1 vial 22 mL
3. Calibrator: Factor value (f) stated on label. Red Cap.	1 vial 150 µL
4. Negative Control: Range Stated on label. Natural cap.	1 vial 150 µL
5. Positive Control: Range Stated on label. Brown cap.	1 vial 150 µL
6. Washing Concentrate (H) 20X:	1 bottle (50 mL)
7. Enzyme Conjugate: Red Color Solution	1 vial (12 mL)
8. TMB Chromogenic Substrate: Amber Bottle	1 vial (12 mL)
9. Stop Solution	1 vial (12 mL)

STORAGE AND STABILITY

1. Store the kit at 2 - 8 °C.
2. Always keep microwells tightly sealed in pouch with desiccants. We recommend you use up all wells within 4 weeks after initial opening of the pouch.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light during storage or usage.

WARNINGS AND PRECAUTIONS

1. Potential biohazardous materials:
The calibrator and controls contain human source components which have been tested and found nonreactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, as there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent, these reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories." 1984

2. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
3. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
4. This product contains components preserved with sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azide. On disposal, flush with a large volume of water.

SPECIMEN COLLECTION AND HANDLING

1. Collect blood specimens and separate the serum.
2. Specimens may be refrigerated at 2 - 8 °C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing of serum sample.

SUMMARY OF ASSAY PROCEDURE

Step	(20-25° C Room temp.)	Volume	Incubation time
1	Sample dilution 1:40 = 5 µl / 200 µl		
2	Diluted sample, calibrator & controls	100 µl	30 minutes
3	Washing buffer (3 times)	350 µl	
4	Enzyme conjugate	100 µl	30 minutes
5	Washing buffer (3 times)	350 µl	
6	TMB Chromogenic Substrate	100 µl	15 minutes
7	Stop solution	100 µl	
8	Reading OD 450 nm		

PREPARATION FOR ASSAY

1. Prepare 1x washing buffer.
Prepare washing buffer by adding distilled or deionized water to 20 x wash concentrate to a final volume of 1 liter.
2. Bring all specimens and kit reagents to room temperature (20-25°C) and gently mix.

ASSAY PROCEDURE

1. Place the desired number of coated strips into the holder.
2. Prepare 1:40 dilutions by adding 5 µl of the samples, negative control, positive control and calibrator to 200 µl of sample diluent. Mix well.
3. Dispense 100 µl of diluted sera, calibrator, and controls into the appropriate wells. For the reagent blank, dispense 100 µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 30 minutes at room temperature.
4. Remove liquid from all wells. Repeat washing three times with washing buffer.
5. Dispense 100 µl of enzyme conjugate to each well and incubate for 30 minutes at room temperature.
6. Remove enzyme conjugate from all wells. Repeat washing three times with washing buffer.
7. Dispense 100 µl of TMB Chromogenic Substrate to each well and incubate for 15 minutes at room temperature.

8. Add 100 µl of Stop solution to stop reaction.
Make sure there are no air bubbles in each well before reading
9. Read O.D. at 450 nm with a microwell reader.

CALCULATION OF RESULTS

1. To obtain Cut off OD value: Multiply the OD of Calibrator by Factor (f) printed on label of Calibrator.
2. Calculate the IgG Index of each determination by dividing the OD values of each sample by obtained OD value of Cut off.

For example:

If Factor (f) value on label = 0.4

This factor (f) is a variable. It is specific for a lot manufactured and printed on label of Calibrator.

Obtained Calibrator O.D. = 1.100

Cut-off O.D. = $1.100 \times 0.4 = 0.44$ (By definition IgG Index = 1)

Patient sample O.D. = 0.580

IgG Index = $0.580 / 0.44 = 1.32$ (Positive result)

Patient sample O.D. = 0.320

IgG Index = $0.320 / 0.44 = 0.73$ (Negative result)

QUALITY CONTROL

The test run may be considered valid provided the following criteria are met:

1. The O.D. value of the reagent blank against air from a microwell reader should be less than 0.150.
2. If the O.D. value of the Calibrator is lower than 0.250, the test is not valid and must be repeated.
3. The HSV 2 IgG Index for Negative and Positive Control should be in the range stated on the labels.

INTERPRETATION

Negative: HSV 2 IgG Index of 0.90 or less are seronegative for IgG antibody to HSV 2.

Equivocal: HSV 2 IgG Index of 0.91 - 0.99 is equivocal. Sample should be retested.

Positive: HSV 2 IgG Index of 1.00 or greater are seropositive.

PERFORMANCE CHARACTERISTICS

Sensitivity and Specificity

Sensitivity, specificity and accuracy were evaluated using a commercial available ELISA kit on 32 specimens. The correlation results are summarized in the following table:

		Reference ELISA		
		N	P	Total
DAI ELISA	N	11 (D)	1 (B)	12
	P	0 (C)	20 (A)	20
	Total	11	21	32

Sensitivity = $A/(A+B) = 20/(20+1) = 95.24\%$

Specificity = $D/(C+D) = 11/11 = 100\%$

Accuracy = $(A+D)/(A+B+C+D) = (20+11)/32 = 96.88\%$

Precision:

The precision of the assay was evaluated by testing three different sera of eight replicates over a period of one week.

The intra-assay and inter-assay C.V. are summarized below:

	Negative	Low positive	Positive
Intra-assay	10.5%	8.7%	7.6%
Inter-assay	11.5%	9.8%	8.6%

LIMITATION OF THE PROCEDURE

1. As with other serological assays, the results of these assays should be used in conjunction with information available from clinical evaluation and other diagnostic procedures.
2. Samples obtained too early during primary infection may not contain detectable antibody.
3. A single serum sample should not be used to aid in the diagnosis of recent infection. Paired samples should be collected and tested simultaneously to look for seroconversion.

REFERENCES

1. Nahmias, A.J., J. Dannenbarger, C. Wickliffe and M. Muther. Clinical aspects of infection with herpes simplex viruses 1 and 2 in the human herpes viruses. An interdisciplinary Perspective (Nahmias, A.J., W.R. Dawdle and R.F. Schinazi eds) New York, Elsevier, pp 3-9, 1981.
2. Vestergaard, B.F., P.C. Grauballe and H. Spanggaard. Titration of herpes simplex virus antibodies in human sera by the enzyme-link immunosorbent assay (ELISA). Acta Pathol. Microbiol. Scand. Sect. B 85:446-448, 1977.
3. Coleman, R.M., L. Pereira, P.D. Bailey, D. Dondero, C. Wickliffe, and A.J. Nahmias. Determination of herpes simplex virus type-specific antibodies by enzyme-linked immunosorbent assay. J. Clin. Microbiol. 18 (1983) 287.

Date Adopted	2017-08-18
REF 1403-11	DAI-HSV2 IgG
EC REP	CEpartner4U, Esdoornlaan 13, 3951DB Maarn. The Netherlands. www.cepartner4u.eu



DIAGNOSTIC AUTOMATION, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, CA 91367

Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision Date: 2017-07