



DIAGNOSTIC AUTOMATION, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, California 91367 USA

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com



See external label



2°C-8°C



96 tests



5142-8

Enzyme Immunoassay for the Quantitative Determination of Mustard in Food

MUSTARD

REF 5142-8

Sensitivity	1 PPM
Recovery (spiked samples)	76-98%
Incubation Time	60 min

GENERAL INFORMATION

Mustard belongs to the Brassica plants. With about 30-35% the fraction of proteins in mustard seed is very high. Some of these proteins are known for being allergenic, such as Sin a 1 and Bra j 1. These proteins are predominantly heat resistant making them stable to different production processes. In addition to brown mustard (*Brassica juncea*) and black mustard (*Brassica nigra*) primarily yellow mustard (*Sinapsis alba*) is used as an ingredient in many foods and food preparations. For mustard allergic persons hidden mustard allergens in food are a critical problem. Already very low amounts of mustard can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, mustard allergic persons must strictly avoid the consumption of mustard or mustard containing food. Cross-contamination, mostly in consequence of the production process, is often noticed. The sau-sage production process is a representative ex-ample. This explains why in many cases the exis-tence of mustard residues in food cannot be ex-cluded. For this reason sensitive detection systems for mustard residues in foodstuffs are required.

The **DAI Mustard ELISA** represents a highly sensitive detection system for yellow mustard and is particularly capable of the quantification of residues in sausage, dressings, soups, cheese and mixed herbs. Due to high cross-reactivity the test is also suitable for the detection of brown mustard and black mustard.

PRINCIPLE OF THE TEST

The **DAI Mustard** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against mustard proteins is bound on the surface of a microtiter plate. Mustard containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against mustard proteins is given into the wells and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of mustard is directly proportional to the colour intensity of the test sample.

PRECAUTIONS

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micropipets, ELISA reader etc.).

HEALTH AND SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

REAGENTS

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are printed on the labels of the bottles and the outer package.

1. Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with anti-mustard antibodies.
2. Mustard Standards (0; 2; 6; 20; 60 ppm of mustard): 5 vials with 1.0 mL each, dyed red, ready-to-use.
3. Conjugate (anti-mustard-peroxidase): 15 mL, dyed red, ready-to-use.
4. Substrate Solution (TMB): 15 mL, ready-to-use.
5. Stop Solution (0.5 M H₂SO₄): 15 mL, ready-to-use.
6. Extraction and sample dilution buffer (Tris): 2 x 120 mL as 10x concentrate, dyed red. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Plastic bag to store unused microtiter strips.
9. Instruction Manual.

Additional Instrumentation and Reagents (not provided)

Instrumentation

- 100 - 1000 µL micropipets
- Volumetric flask
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)

Reagents

- double distilled water

SAMPLE PREPARATION

Due to high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be cleaned thoroughly before and after each sample. Mustard proteins could adhere to different surfaces. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for all kinds of samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill, etc.
2. 1 g of the homogenized mixture is suspended in 20 mL of pre-diluted extraction buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2500 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated.
4. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the pre-diluted extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

PROCEDURE

The washing solution is supplied as 10x concentrate and has to be diluted 1+9 with double distilled water before use. In any case the ready-to-use standards provided should be determined twofold. When samples in great quantities are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 µL ready-to-use standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 µL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.
5. Pipet 100 µL of conjugate (anti-mustard-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 µL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 µL of stop solution (0.5 M H₂SO₄) into each well. The blue colour will turn yellow upon addition.

11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

CALCULATION OF RESULTS

The ready-to-use standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppm on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of mustard in ppm from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

TYPICAL STANDARD VALUES

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 60 ppm standard. These values are only an example and should not be used instead of the standard curve which has to be measured in each new test.

Mustard (ppm)	% binding of 60 ppm
60	100
20	50
6	21
2	12
0	7

PERFORMANCE

Sensitivity

The limit of detection (LOD) of the DAI Mustard test is 1 ppm.

The limit of quantification (LOQ) of the DAI Mustard test is 2 ppm.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

Cross-reactivity

For the following foods no cross-reactivity could be detected:

Milk	Corn	Radish
Egg	Buckwheat	Cabbage
Pork	Soy	Pepper
Beef	Sesame	Curcuma
Chicken	Fish (Cod)	Cayenne
Wheat	Pea	Clove
Rye	Bean	Nutmeg
Oats	Carrot	Cinnamon
Barley	Leek	Dill
Rice	Celery	Thyme

The following cross reactions were determined:

Horseradish	0.0007%
Garden cress	0.0009%
Garden cress (seed)	1.5%
Rape (seed)	15.5%
Radish (seed)	31.2%
Cabbage (seed)	29.2%
Brown mustard (seed)	26.5%
Black mustard (seed)	32.5%

Precision

Intra-assay Precision	8%
Inter-assay Precision	12%

Linearity

The serial dilution of spiked samples (sausage, salad dressing, instant soup, canned soup, cheese, mixed herbs) resulted in a dilution linearity of 81% - 117%.

Recovery

Mean recovery was determined by spiking samples with different amounts of mustard:

Sausage	98%
Salad dressing	76%
Instant soup	80%
Canned soup	96%
Cheese	89%
Mixed herbs	78%

REFERENCE

1. Lee PW, et al. (2008) – Sandwich enzyme-linked immunosorbent assay (ELISA) for detection of mustard in food. J Food Sci, 73(4):62-68
2. Koppelman SJ, et al. (2007) – Development of an enzyme-linked immunosorbent assay method to detect mustard protein in mustard seed oil. J Food Prot, 70(1):179-83

3. Shim YY, et al. (2008) - Quantitative detection of allergenic protein Sin a 1 from yellow mustard (sinapsis alba L.) seeds using enzyme-linked immunosorbent assay. J Agric Food Chem, 56(4):1184-92
4. Lee PW, et al. (2009) – Detection of mustard, egg, milk and gluten in salad dressing using enzyme-linked immunosorbent assays (ELISAs). J Food Sci, 74(5):46-50
5. Palomares, et al. (2005) – Isolation and identification of an 11S globulin as a new major allergen in mustard seeds. Ann Allergy Asthma Immunol, 94(5):586-92

Date Adopted	2015-10-19
REF 5142-8	DA-Mustard



DIAGNOSTIC AUTOMATION, INC.

21250 Califa Street, Suite 102 and 116, Woodland Hills, California 91367 USA

Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision Date: 2010-04-19