

**OneStep
Methadone
RapiCard™ InstaTest**

Cat # 121005-1



Sensitivity	300 ng/ml
--------------------	------------------

INTENDED USE

The Cortez Diagnostics, Inc. OneStep MTD RapiCard™ InstaTest is an immunochromatography based one step in vitro test.

SUMMARY AND EXPLANATION

Methadone is a synthetic opioid, clinically available. It is used clinically for the treatment of severe pain and in maintenance programs for morphine and heroine addicts.

Methadone acts on the central nervous and cardiovascular systems to produce respiratory and circulatory depression. Methadone also produces miosis and increases the tone of smooth muscle in the lower gastrointestinal tract while decreasing the amplitude of contractions. Acute higher doses induce analgesia, sedation, respiratory depression and coma.

After methadone administration, the major urinary excretion products are methadone and its metabolites, EDDP and EMDP. Large individual variations in the urine excretion of methadone are output of methadone

TEST PRINCIPLE

The Cortez Diagnostics, Inc. OneStep MTD RapiCard™ InstaTest is based on the principle of specific immunochemical reaction between antibodies and antigens to analyze particular compounds in human

urine specimen. The assay relies on the competition for binding mouse anti-methadone monoclonal antibody between methadone-protein conjugate and free drug which may be present in the urine specimen being tested.

When methadone is present in the urine specimen, it competes with methadone-protein conjugate for the limited amount of monoclonal antibody-colloidal gold conjugate. When the amount of methadone is equal or more than the cut-off, it will prevent the binding of methadone-protein conjugate to the antibody. Therefore, a positive urine specimen will not show a colored band on the test line zone, indicating a positive result, while the presence of a colored band indicates a negative result.

A control line composed of Goat anti-Mouse IgG antibody is present in the test window to work as procedural control. This colored band should always appear on the control line zone if the test device is stored in good condition and the test is performed appropriately.

SPECIMEN COLLECTION AND PREPARATION

It is required that approximately 150 µl of sample for each test. Fresh urine specimens do not need any special handling or treatment. Specimens should be collected in a clean, dry, plastic or glass container. If the assay is not performed immediately, urine specimen may be refrigerated at 2-8 C or frozen up to 7 days. Specimens should be thawed and brought to room temperature before testing. Urine specimens exhibiting a large amount of precipitate or turbidity should be centrifuged or allowed to settle before testing.

MATERIALS AND COMPONENTS

Materials provided with the test kits

1. Instructions for use.
2. Cortez MTD Test device. The amount of each coated antigen and/or antibody on the strip is less than 1.0 mg for antigen conjugate and is less than 1.0 mg for goat anti-mouse IgG antibody.
3. Test zone: contains Methadone bovine protein antigen conjugates.
4. Control zone: contains Goat anti-mouse IgG antibody.
5. Conjugate pad: contains about 2 µg of colloidal gold-mouse anti-methadone monoclonal antibody conjugate.

Materials required but not provided

1. Urine collection container.
2. Timer or clock.

ASSAY PROCEDURE

1. Bring all materials and specimens to room temperature.
2. Remove the test card from the sealed foil pouch.
3. Place the transfer pipette in the specimen and depress the bulb to withdraw a sample.
4. Hold the pipette in a vertical position over the sample well of the test card and deliver 2-3 drops (100-150 µl) of sample into the well.
5. Read the results at 5 minutes after adding the sample.
Do not interpret the result after 5 minutes.

RESULTS

- **Negative:** Two colored bands form. The appearance of two colored bands, one in test line zone and the other in control line zone, indicates negative results. The negative result indicates that the methadone concentration in the specimen is either zero or less than cut-off level.
- **Positive:** One colored band forms. One colored band appears in control line zone. No colored band is found in test line zone. This is an indication that the methadone level in the specimen is above the cut-off level.
- **Invalid:** If there are no colored bands in control line zone, the test result is invalid. Retest the sample with a new device.

Note: A *borderline (+/-) in test line zone be considered negative result.*

PERFORMANCE CHARACTERISTICS

A. Accuracy

The accuracy of the Cortez OneStep MTD RapiCard™ InstaTest was evaluated in comparison to GC/MS at a cut-off of 300 ng/ml of methadone. One hundred and nineteen urine specimens with confirmed methadone concentrations were evaluated in this study. Borderline readings were recorded as negative. The results are summarized and presented below:

Cortez MTD Test	GC/MS Negative (less than -25% cut off)	(-)			(+) GC/MS Positive (greater than +25% cut off)			Percent agreement with GC/MS
		Near cutoff negative (between -25% and c/o)	Near cutoff positive (between c/o and +25%)					
Positive	0	0	9	49	100			
Negative	50	9	2	0	96.7			
Total	50	9	11	49				

Positive % agreement: 96.7, Negative % agreement: 100.

Two specimens were found discrepant between the Cortez MTD and GC/MS method. When compared those data, 100% (2 out of 2) of the discrepancy specimens were found between cut-off and +25% cut-off concentration (300 – 375 ng/ml).

B. Sensitivity

The cut-off concentration (sensitivity level) of Cortez MTD test is determined to be 300ng/ml.

C. Precision

The precision study was performed by three individuals observing the test result to determine the random error of visual interpretation. The test results were found to have no significant differences between the three observers.

Device	Control Con. ng/ml	No. of Tested	No. of positive			No. of borderline #			No. of negative		
			1*	2*	3*	1*	2*	3*	1*	2*	3*
	150	40							40	40	40
	225	40				21	28	23	19	12	17
MTD	300	40	23	27	23	17	13	17			
	375	40	40	40	40						
	450	40	40	40	40						

1. Interference Testing

The Cortez Diagnostics, Inc. OneStep MTD RapiCard™ InstaTest performance at cut-off level is not affected when pH and Specific Gravity ranges of urine specimen are at 4.5 to 8.0 and 1.005 to 1.035.

The following substances were tested and confirmed not to interfere with Cortez MTD test at the listed concentrations.

Glucose	2000 mg/dl,
Human albumin	2000 mg/dl
Human hemoglobin	10 mg/dl,
Urea	4000 mg/dl
Uric acid	10 mg/dl

2. Specificity

The following table lists compounds that are detected by Cortez MTD test which produced positive results when tested at levels equal or greater than the concentrations listed below:

Compounds	Cut-off (ng/ml)	Cross reactivity (%)
Methadone	300	100
Doxylamine	50,000	0.6
Norpropoxyphene	1,000	30

The following compounds show negative results at concentration up to 100 µg/ml unless specified.

Acetaminophen	Acetylsalicylic acid	Amikacin
Amobarbital	Amphetamine	Arterenol
Ascorbic acid	Atropine	Benzoic acid
Butabartital	Caffeine	Camphor
Chloroquine	Cocaine	Cortisone
Dextromethorphan	Diazepam	Digitoxin
Diphenhydramine	Ecgonine	Ecgonine methyl ester
Epinephrine	Gentisic acid	Guaiacol glycer ester
Homatrophine	Hydrochlorothiazide	Ibuprofen
Isoproterenol	Ketamine	Lidocaine
Methamphetamine	Meperidine	Methaqualone
Morphine	Neomycin	Niacinamide
Perphenazine	Penicillin G	Phencyclidine
Phenylethylamine-α	Phenylpropanolamine	Promethazine
Quinine antidine	Salicylic acid	Secobarbital
Tetrahydrozoline	Theophylline	11-nor-Δ ⁸ -THC-9-COOH(10 µg/ml)
11-nor-Δ ⁹ -THC-9-COOH(10 µg/ml)	Tryptophan	Thioridazine
Tyramine		Propoxyphene

Amitriptyline	Histamine
Aspartame	Imipramine 11-nor-Δ ⁹ -THC-9-COOH(10 µg/ml)
Benzoylcegonine	3, 4±MDA
Chlopheniramine	Methylphenidate
Deoxyephedrine	Oxazepam
Digoxin	Phenobarbital
Ephedrine	Pseudoephedrin
Histamine	Tetracycline
	Trifluoperazine

QUALITY CONTROL

Good Laboratory Practice recommends the daily use of control materials to validate the reliability of device. Control materials should be assayed as clinical specimen and challenging to the assay cutoff concentration, e.g., 25% above and below cutoff concentration. If control values do not fall within establish range, assay results are invalid. Control materials which are not provided with this test kit are commercially available.

The Cortez Drugs of Abuse Test provides a built-in process control with a different antigen/antibody reaction at the control region (C). This control line should always appear regardless the presence of drug or metabolite. If the control line does not appear, the test device should be discarded and the obtained result is invalid. The presence of this control band in the control region serve as 1) verification that sufficient volume is added, 2) that proper flow is obtained.

The control line does not appear, the test should be discarded and the obtained result is invalid.

You should always follow local, state and federal guidelines for running QC.

LIMITATION OF PROCEDURE


The assay is designed for use with human urine only. This test has not been evaluated in a point of care setting. A positive result with any of the tests indicates only the presence of a drug/metabolite and does not indicate or measure intoxication. There is a possibility that technical or procedural error as well other substances in certain foods and medicines may interfere with the test and cause false results. Please refer "SPECIFICITY" section for lists of substances that will produce either positive results, or that do not interfere with test performance. If a drug/metabolite is found present in the urine specimen, the assay does not indicate frequency of drug use or distinguish between drug of abuse and certain foods and medicines.

EXPECTED VALUES

The Cortez OneStep MTD RapiCard™ InstaTest is a qualitative assay. It identifies methadone in human urine at a concentration of 300 ng/ml or higher. The concentration of the methadone cannot be determined by this assay. The test is intended to distinguish negative result from presumptive positive result. All positive results must be confirmed using an alternate method, preferably GC/MS.

PRECAUTION

1. For in vitro diagnostic and forensic use only.
2. Do not use the product beyond the expiration date.
3. Handle all specimens as potentially infectious.
4. Do not open foil pouch until it is ready to be tested.
5. Use a new urine specimen cup for each sample to avoid cross contamination.
6. The test device should be stored at 2 to 30°C and will be effective until the expiration date.

<p>ISO 13485 ISO 9001</p>  <p>Diagnostic Automation/ Cortez Diagnostics, Inc. 23961 Craftsman Road, Suite E/F, Calabasas, California 91302 USA</p>			
Date Adopted	Cat # 121005-1		
2008-08-09	<p>CORTEZ- OneStep Methadone RapiCard™ InstaTest -2012</p> <p>CEpartner4U, Esdoornlaan 13, 3951DB Maarn. The Netherlands.</p> <p>www.cepartner4u.eu</p>		
<table border="1" style="display: inline-table;"> <tr> <td>EC</td> <td>REP</td> </tr> </table>	EC	REP	
EC	REP		
Revision Date: 12-20-2013			